

CORRIGENDUM/ADDENDUM

Tender No. N/PUS/P/2020/06/13 for the ONLINE TENDER FOR procurement of REFRIGERATED CENTRIFUGE FOR AES LAB

Reference: 1) Date of Issue of Tender Document: 07.07.2021

2) Date of Pre-bid Meeting : 13.07.2021

Please refer the pre-bid meeting held on above cited date for the procurement of above equipments.

On the recommendation of User as well as Technical Evaluation Committee, the following corrigendum/addendum is proposed as under:

Clause No.	Present Para/Specification	Proposal Corrigendum/Addendum (to be read)
	We include the following point : 1. Minimum Temperature 2. Speed with max. 14,000 rpm / 20,830 xg RCF or better 3. Certification/Compliance : European CE Certified , UL Listed, CSA Certified, IVD Compliant 4. Fixed Rotor Angel: 6 x 85 mL capacity to accommodate rotor accommodate 15 mL & 50 mL conical tubes thru suitable adaptors (6 each) max. RPM 12000/18000 xg or better	1. The rotor must maintain 4°C at max. Speed. 2. Speed with max. 14,000 rpm / to 20,000xg or better <i>Tripti</i> 3. Remove the IVD Certificate. 4. Fixed Rotor Angel: 6 x 50 mL capacity (or better) to accommodate 15 mL & 50 mL conical tubes thru suitable adaptors (6 each) max. RPM 12000/18000 xg or better.

The said tender may be read amended to this extent. The other terms & condition of tender are remain unchanged.

Tripti
20/07/21
(Dr. Tripti Aggarwal)
Assistant Professor

S
20/07/2021
(Dr. S.Chakkaravarthy)
Assistant Professor

Neeraj
20/7/21
(Dr. Neeraj)
Associate Professor